


APPROVED

Production Director
Underground Mining
LLC "RMG Auramine"



" 28 " 07 2026

V. Abashmadze

AGREED

Chief Geologist
LLC "RMG Auramine"



" 28 " 07 2026

N. Tatuashvili

TECHNICAL SPECIFICATION

Procurement of an Integrated Jaw Crusher with Automatic Rotary Sample Divider

1. Procurement Object

This technical specification defines the minimum technical requirements for an integrated jaw crusher with an automatic rotary sample divider intended for the Geological Sample Preparation Laboratory of the Berthakari Deposit of LLC "RMG Auramine". The equipment shall be a single factory-built unit integrating the jaw crusher and the automatic rotary sample divider into one system.

2. Scope of Application

The equipment is intended for primary crushing of geological samples, ores and rocks, automatic representative sample division of crushed material, and preparation of representative laboratory samples for analysis.

3. Scope of Supply

1. Integrated jaw crusher with automatic rotary sample divider.
2. All major mechanical and electrical components required for operation.

No.	Parameter	Requirement
1	Equipment type	Integrated jaw crusher with automatic rotary sample divider

2	Crusher capacity	Approx. 2–2.5 t/h
3	Maximum feed size	150–170 mm
4	Output size	5–30 mm
5	Discharge gap adjustment	Required
6	Construction	Heavy-duty integrated factory-built construction
7	Sample divider	Integrated automatic rotary sample divider with independent electric motor
8	Power supply	380 V / 50 Hz / Three-phase
9	Construction and hopper material	Sample receiving hoppers shall be made of corrosion-resistant stainless steel. The machine structure shall be coated with high-quality industrial paint.
10	Safety system	All rotating and moving parts shall be protected by safety guards.
11	Warranty	Minimum 12 months
12	Compliance	CE Certificate or equivalent conformity document
13	Technical documentation	Technical passport, operating manual and certificates of conformity
14	Operating principle	Crushed material shall be automatically transferred from the crusher to the integrated rotary sample divider without manual handling.
15	Condition	New, unused, factory manufactured

Prepared by
Head of Sample Preparation Laboratory
LLC "RMG Auramine"

ს. ჯგერა Avtandil Vardzelashvili